


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVRL-1TT7B-N

S.N.: 2710

Date: JAN 15 86




- a) Output signal amplitude:
A) $-200V$
B) $-200V$
- b) Pulse width:
A) 10 TO 100 NSEC
B) 100 NSEC TO 100 USEC
- c) Rise time:
A) $\leq 5 NSEC$
B) $\leq 5 NSEC$
- d) Fall time:
A) $\leq 5 NSEC$
B) $\leq 5 NSEC$
- e) PRF:
0 TO 50 KHz

A out 50 VOLT/DIV
50 NSEC/DIV

- f) Jitter, stability:
OK
- g) Prime power:

$120/240V$
 $50-60Hz$



B out 50 VOLT/DIV
20 USEC/DIV

- h) DELAY
(SYNC TO OUT)
0 TO 50 NSEC
(IN TO SYNC)
100 NSEC

[Handwritten signature]