


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMK-F*

S.N.: *2758*

Date: *SEPT 25 '85*



- a) Output signal amplitude:
> +5 VOLTS
- b) Pulse width:
4 TO 100 NSEC
- c) Rise time: *(SEE FILE 1, 2)*
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
100 NSEC/DIV

+15 ± 1.0 VOLT
250 mA (MAX)
[Signature]