


PULSE GENERATOR

PERFORMANCE CHECK


Model: APRL-1777A-N

S.N.: 2759

Date: NOV 15 85



A out 50 V/DIV / DIV
100 NS/DIV



B out 50 V/DIV / DIV
200 NS/DIV

- a) Output signal amplitude:
A) -200 V } $R_L \geq 1K$
B) -200 V }
- b) Pulse width:
A) 10 TO 100 NS
B) 100 TO 800 NS
- c) Rise time:
A) < 15 NS
B) < 5 NS
- d) Fall time:
A) ≤ 5 NS
B) ≤ 5 NS
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50-60 Hz

h) 20249
100 NS TO 65 us