


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANMP-2-C-0T*

S.N.: *2767*

Date: *OCT 1 1996 (REPAIRS & RETAIL)*



a) Output signal amplitude:

*0 TO +10 VOLTS*

b) Pulse width:

*5 TO 10 NS*

c) Rise time:

*≤ 100 PS*

d) Fall time:

*≤ 135 PS*

e) PRF:

*0 TO 1 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120/240V  
50-60 Hz.*

*(A) 10V/DIV, 5NS/DIV  
PRF = 100 KHz*

*(B) RISE TIME, 100 PS/DIV  
10V/DIV, PRF = 100 KHz*

*(C) FALL TIME, 100 PS/DIV  
10V/DIV, PRF = 100 KHz*