



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PAAMP-3-C-PN*  
S.N.: *2308*  
Date: *JAN 14 '86*



Ⓐ 10 VOLTS/DIV  
10 NSEC/DIV  
1 MHz



Ⓑ AS Ⓐ BUT  
1.0 NSEC/DIV  
(RISE TIME)

- a) Output signal amplitude:  
*0 TO ± 20 VOLTS*
- b) Pulse width:  
*5 TO 100 NSEC*
- c) Rise time:  
*≤ 200 PSEC*
- d) Fall time:  
*≤ 300 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK.*
- g) Prime power:  
*120/240 V*  
*50 - 60 MHz*