


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-P-MM1*

S.N.: *2814*

Date: *OCT 30 85*



- b) Output signal amplitude:  
*± 15 VOLTS*
- c) Pulse width:  
*4 TO 100 NS (SEE FIG 1)*
- d) Rise time:  
*≤ 100 PS*
- e) Fall time:  
*≤ 2 NS*
- f) PRF: *MIN PW: 1 MHz* } *SEE FIG 1+2*  
*MAX PW: 20 kHz*
- g) Jitter, stability:  
*OK*
- h) Prime power:

*10 VOLTS/DIV*  
*1.0 NS/DIV*  
*500 kHz*

*+15 ± 1.0 VOLTS*  
*200 mA (MAX)*  
*[Signature]*