


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-P-MMI.*

S.N.: *2815*

Date: *OCT 30 85*



-) Output signal amplitude:
> +15 VOLTS.
-) Pulse width:
4 TO 100 NSEC.
-) Rise time: *(SEE FIG 1)*
≤ 100 PSEC
-) Fall time:
≤ 2 NSEC
-) PRF: *MIN PW: 0 TO 100 KHZ) SEE*
MAX PW: 0 TO 20 KHZ) FIG 1+2.
-) Jitter, stability:
OK
-) Prime power:

10 VOLTS/DIV
1.0 NSEC/DIV
500 KHZ.

+15 ± 1.0 VOLTS.
200 mA (MAX.)