


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANI-V-PS-PN-05-M

S.N.: 2839

Date: FEB 4 '86



-) Output signal amplitude:
0 TO ± 50 VOLTS
-) Pulse width:
3 TO 100 NSEC
-) Rise time:
≤ 350 PSEC
-) Fall time:
≤ 2 NSEC
-) PRF:
0 TO 20 KHz
-) Jitter, stability:
OK
-) Prime power:

≈ 32 VOLTS/DIV
5.0 NSEC/DIV
20 KHz
 $R_L = 50 \Omega$

120/240 V
50-60 Hz