


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-TN-1
S.N.: 2841
Date: DEC 16 85



-) Output signal amplitude:
 - a) - 230 VOLTS
 - b) +30 VOLTS
-) Pulse width:
0 TO 100 NSEC.
-) Rise time:
≤ 2 NSEC.
-) Fall time:
≤ 3 NSEC.
-) PRF:
0 TO 5 KHz
-) Jitter, stability:
OK
-) Prime power:

100 NSEC/DIV
10 NSEC/DIV
1 KHz.

120/240V
50-60 Hz.
h) t_p (PROPAGATION DELAY)
≤ 100 NSEC

[Handwritten signature]