


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVH-1P*
S.N.: *2858*
Date: *DEC 8 '85*



- a) Output signal amplitude:
0 to +30 VOLTS
- b) Pulse width:
1.0 NS ± 2%
- c) Rise time:
NP
- d) Fall time:
NA
- e) PRF:
0 to 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
1.0 NS/DIV
500 KHz

+15 ± 1.0 VOLT
150 mA (MAX)
[Signature]