


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APRL-ITTSB-N*

S.N.: *2859*

Date: *FEB 3 86*



100 nsec/div



B) 20 us/div

- a) Output signal amplitude:
A out): -250 VOLTS
B out): -250 VOLTS
- b) Pulse width:
A out): 20 TO 100 nsec
B out): 100 nsec TO 5 msec
- c) Rise time:
≤ 5 nsec
- d) Fall time:
≤ 5 nsec
- e) PRF:
0 TO 30 Hz
- f) Jitter, stability:
- g) Prime power:
OK
120/240 V
50-60 Hz
- h) DELAY:
A out): 30 nsec TO 2 us
B out): 100 nsec TO 2 us

[Signature]