


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-P-MMM1*

S.N.: *2863*

Date: *NOV. 15 85*



- ) Output signal amplitude:  
*> +15 VOLTS.*
- ) Pulse width:  
*4 TO 100 NSEC.*
- ) Rise time:  
*≤ 100 PSEC*
- ) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*MIN PW : 0 TO 1 MHz*  
*MAX PW : 0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*500 KHz*

*+15 ± 1.0 VOLTS.*  
*200 mA (MAX)*