


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MM-2-C-PN-D-M-OT*

S.N.: *2876*

Date: *FEB 19 86*



- a) Output signal amplitude:  
*0 TO ±15 VOLTS*
- b) Pulse width:  
*0.2 TO 2.0 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV  
1.0 NSEC/DIV  
10 MHz*

*120/240 V,  
50 - 60 Hz*