


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-HV1-C-PW-25-M*

S.N.: *2379*

Date: *MAR 20 86*



- a) Output signal amplitude:  
*0 TO ±50 VOLTS*
- b) Pulse width:  
*2 TO 100 NSEC*
- c) Rise time:  
*≤ 250 PSEC.*
- d) Fall time:  
*< 2 NSEC*
- e) PRF:  
*0 to 20 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 32 VOLTS/PW  
5.0 NSEC/PW  
20 KHz.*

*120/240 V  
50-60 Hz.*