


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-3A-C-P-LE1*

S.N.: *2886*

Date: *MAR 3 86*



- a) Output signal amplitude: *+25 to +75 Amps*
- b) Pulse width: *5 nsec*
(100% power)
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 5 kHz*
- f) Jitter, stability: *OK*
- g) Prime power:

10 nsec/div
50 Amps/div
5 kHz
R_i = 2.0 Ω

120/240 V,
50-60 Hz
[Signature]