


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-PS-1A-CA1*

S.N.: *2891*

Date: *JAN 8 '96*



- a) Output signal amplitude:
≥ +2.5 VOLTS
- b) Pulse width:
2 TO 4 NSEC
- c) Rise time:
< 500 PSEC
- d) Fall time:
< 500 PSEC
- e) PRF:
5 TO 22 MHz
(OPTIMIZED FOR 22 MHz)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 Hz

10 VOLTS/DIV
2 NSEC/DIV
22 MHz

[Signature]