


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMN-1-P*

S.N.: *2905*

Date: *FEB 11 86*



- 1) Output signal amplitude:
0 TO +5 VOLTS
- 2) Pulse width:
0.2 TO 5 nsec.
- 3) Rise time:
≤ 100 psec.
- 4) Fall time:
≤ 200 psec.
- 5) PRF:
0 TO 50 MHz.
- 6) Jitter, stability:
OK
- 7) Prime power:

*5 VOLTS/DIV
1.0 nsec/DIV
10 MHz.*

*+24 ± 1.0 VOLTS
400 mA (MAX)*