


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AM-1-C-P-PN-0T*
S.N.: *2908*
Date: *FEB 18 86*



- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
0.2 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 NSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 NSEC/DIV
10 MHz

120/240 VOLTS
50-60 Hz