


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-D-PN-ID-M-OT*

S.N.: *2943*

Date: *MARK 12 86*



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0 TO 6 NSER
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz

[Signature]

5 VOLTS/DIV
1.0 NSER/DIV
12 MHz