


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MR-B3-EP-P*

S.N.: *2945*

Date: *APRIL 2 26*



- a) Output signal amplitude:  
*0 TO +100 VOLTS*
- b) Pulse width:  
*20 TO 200 NSEC*
- c) Rise time:  
*53 NSEC*
- d) Fall time:  
*53 NSEC*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 VOLTS/DIV*  
*50 NSEC/DIV*  
*20 KHz.*  
 *$R_L = 50 \Omega$*

*+24 ± 1.0 VOLTS*  
*250 mA (MAX)*

*[Signature]*