


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *2956*
Date: *FEB 5 '86*



- a) Output signal amplitude:
+25 TO +100 AMP
- b) Pulse width:
25 ± 5 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 60 AMP/DIV
10 NSEC/DIV
1 KHz

+15 ± 1.0 VOLT.
200 mA (MAX)