

PULSE GENERATOR  
PERFORMANCE CHECK

Model:



APRL-7T5C-N

S.N.:

2963

Date:

MAR 7 81



- a) Output signal amplitude:
  - A) - 250 VOLTS
  - B) - 250 VOLTS
- b) Pulse width:
  - A) 20 TO 100 NSEC
  - B) 100 NSEC TO 20 USEC
- c) Rise time:
  - ↳ 10 NSEC
- d) Fall time:
  - ↳ 10 NSEC
- e) PRF:
  - 0 TO 50 MB
- f) Jitter, stability:
  - OK
- g) Prime power:

120/240 V,  
50 - 60 Hz

- h) DELTA:
  - A) 30 NSEC TO 1 MSEC
  - B) 100 NSEC TO 1 MSEC