


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PR-3-PW-C-7*

S.N.: *2965*

Date: *MAR 11 86*



- a) Output signal amplitude:
0 TO + 200 VOLTS
- b) Pulse width:
0.1 TO 100 USEC.
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC.
- e) PRF:
MAX PW: 0 TO 50 Hz
MIN PW: 0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:

50 VOLTS/DIV
20 USEC/DIV
R_L = 50 Ω
PRF = 50 Hz

120/240 V
50-60 Hz

[Signature]