


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-2-C-N-PN*

S.N.: *2973*

Date: *FEB 27 86*



- a) Output signal amplitude:  
*0 TO 100 VOLTS*
- b) Pulse width:  
*5 TO 50 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*635 PSEC*
- e) PRF:  
*0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS / DIV*  
*2.0 NSEC / DIV*  
*50 KHZ*

*120 / 240 V*  
*50 - 60 KHZ*