


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PAH-5-1-0-P*  
S.N.: *2983*  
Date: *MAR 13 1986*



- a) Output signal amplitude:  
*0 to +10 VOLTS*
- b) Pulse width:  
*≤ 130 PSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 to 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*500 PSEC/DIV*  
*10 VOLTS/DIV*  
*100 KHz*

*120/240 V*  
*50-60 Hz*  
*[Signature]*