


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-TN-1*

S.N.: *2989*

Date: *FEB 14 86*



- a) Output signal amplitude:
  - a) - 230 VOLTS*
  - b) + 30 V OFFSET*
- b) Pulse width:  
*5 TO 100 NSEC.*
- c) Rise time:  
*≤ 2 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*100 VOLTS/DIV*

*10 NSEC/DIV*

*120/240 V*  
*50-60 HB.*

- h) PROPAGATION DELAY:  
*≤ 100 NSEC.*