


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVOM 4-2-PS-P-PN

S.N.: 2992

Date: APRIL 23 1966



- a) Output signal amplitude:
0 TO ± 30 VOLTS
- b) Pulse width:
 ≤ 400 PSEC.
- c) Rise time:
2 NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz

[Signature]

≈ 10 VOLTS/DIV

1.0 NSEC/DIV

5.0 MHz