


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14 MAZ P*

S.N.: *2995*

Date: *MAY 28 1966*



- a) Output signal amplitude:
≥ +60 VOLTS
- b) Pulse width:
≤ 2 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK.
- g) Prime power:

*32 VOLTS/DIV.
2.0 NSEC/DIV.
500 KHz*

*+135 ± 5 VOLTS
5.0 mA (MAX)*