


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-M-1-P-7N*

S.N.: *3005*

Date: *APRIL 21 '86*



*10 VOLTS/DIV
1.0 NSEC/DIV
1.0 MHz*

- a) Output signal amplitude:
0 TO ±10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 ± 1.0 VOLT
350 mA (MAX)*