


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PWC-N*

S.N.: *3015*

Date: *MAY 8 86*



- a) Output signal amplitude:
0 TO -200 V.
- b) Pulse width:
0.1 TO 100 USEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
MIN PW: 0 TO 1 KHZ
- f) Jitter, stability:
MIN PW: 0 TO 50 HZ.
OK
- g) Prime power:

120/240 V
50-60 HZ

50 VOLTS/DIV
10 USEC/DIV
50 HZ.