


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PL-TN-1*

S.N.: *3057*

Date: *MAR 14 86*



- a) Output signal amplitude:  
*-250 V & +30 V OUT*
- b) Pulse width: *OFFSET*  
*5 TO 100 NSER*
- c) Rise time:  
*≤ 2 NSER*
- d) Fall time:  
*≤ 3 NSER*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120 / 240 V*  
*5.0 - 60 Hz*

*100 VOLTS / DIV*  
*10 NSER / DIV*

- h) *t<sub>p</sub>: ≤ 100 NSER*