


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-7B-SNL1*

S.N.: *3063*

Date: *JUNE 4 26*



- a) Output signal amplitude: *0 TO +30 AMPS TO*
- b) Pulse width: *1  $\mu$*   
*0.5 TO 50  $\mu$ SEC.*
- c) Rise time:  *$\leq$  40 NSEC*
- d) Fall time:  *$\leq$  40 NSEC*
- e) PRF: *MAX PW: 0 to 50 Hz.*  
*PW  $\leq$  20  $\mu$ sec: 0 to 300 Hz.*
- f) Jitter, stability: *OK*

*10 VOLTS/DIV*  
*10  $\mu$ SEC/DIV*  
*50 Hz.*  
*R<sub>L</sub> = 1  $\Omega$ .*

g) Prime power: *120/240 V*  
*50-60 Hz.*