


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PV0-2.4-C-05-M-P*

S.N.: *3125*

Date: *July 14 86*



- a) Output signal amplitude:
0 to +10 Amps
- b) Pulse width:
2 to 50 nsec.
- c) Rise time:
≤ 1 nsec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 to 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz

AA

7 Amps/div
5 nsec/div
10 kHz