


PULSE GENERATOR
PERFORMANCE CHECK


Model: PAI-V-144 P-6A

S.N.: 3144

Date: JUNY 23 86



Ⓐ = 32 VOLTS/DIV
10 NSEC/DIV
10 KHz



Ⓑ AS Ⓐ BUT
1 NSEC/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
2 TO 100 NSEC
- c) Rise time:
≤ 250 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
250 mA (MAX)