


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PK-3-PW-C-CA-N*

S.N.: *3156*

Date: *MAY 9 86*



- a) Output signal amplitude:
0 TO -200 VOLTS
- b) Pulse width:
0.1 TO 100 USEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
MIN PW: 0 TO 1 KHZ
MAX PW: 0 TO 50 MHZ
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 HZ

50 VOLTS/DIV

10 USEC/DIV

R_L = 50 Ω

50 HZ