

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVR-1-PW-PS-P-CA

S.N.:

3162

Date:

MAY 22 86



a) Output signal amplitude:

0 to +200 VOLTS

b) Pulse width:

75 NSEC TO 5.0 USEC.

c) Rise time:

≤ 10 NSEC

d) Fall time:

≤ 10 NSEC

e) PRF:

0 TO 1.0 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240V

50-60 Hz

[Signature]

50 VOLTS/DIV

1.0 USEC/DIV

500 Hz.