


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMP-92-15-C-N*

S.N.: *3172*

Date: *JUL 7 16 86*



Ⓐ *5 VOLT/DIV*
1.0 NSEC/DIV
100 KHz



Ⓑ *AS Ⓐ BUT*
100 PSEC/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO -5 VOLTS
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz