


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ANN-1-C-P*

S.N.: *3207*

Date: *APR 19 86*



Ⓐ *1.0 NSEC/DIV*
5 VOLTS/DIV
≈ 100 MHz.



Ⓑ *1.0 NSEC/DIV*
5 VOLTS/DIV
≈ 250 MHz.

- a) Output signal amplitude:
0.70 + 5 VOLTS
- b) Pulse width:
0.2 DIV 1.0 NSEC
- c) Rise time:
600 PSEC
- d) Fall time:
≤ 100 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz