


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMN-1-C-P-PN-M-0T*


S.N.: *3208*

Date: *APR 15 61*



- a) Output signal amplitude:  
*0 TO ± 5 VOLTS*
- b) Pulse width:  
*0.2 TO 6 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*O.K.*
- g) Prime power:  
*120/240 V*  
*50-60 Hz.*

① *5 VOLTS/DIV*  
*10 NSEC/DIV*  
*(≈ 50 MHz)*



② *5 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*(≈ 14 MHz)*

*[Signature]*