



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PAK-AU-P*  
S.N.: *3229*  
Date: *AUG 13 86*



Ⓐ 10 VOLTS/DIV  
10 nSEC/DIV  
50 KHz



Ⓑ AS Ⓐ BUT  
1.0 nSEC/DIV  
(RISE TIME)

- a) Output signal amplitude:  
*0 TO +15 VOLTS*
  - b) Pulse width:  
*2 TO 50 nSEC*
  - c) Rise time:  
*≤ 100 pSEC*
  - d) Fall time:  
*≤ 2 nSEC*
  - e) PRF:  
*0 TO 50 KHz*
  - f) Jitter, stability:  
*OK*
  - g) Prime power:  
*+15 ± 1.0 VOLTS*  
*300 mA (MAX)*
- [Signature]*