


PULSE GENERATOR
PERFORMANCE CHECK


Model: AVRL-ITTB-B-N

S.N.: 3231

Date: SEPT 30 86



A PULSE 50 VOLTS/DIV
100 NSEC/DIV



B PULSE 50 VOLTS/DIV
100 NSEC/DIV

- a) Output signal amplitude:
 - 200 VOLTS
- b) Pulse width:
 10 TO 100 NSEC
- c) Rise time:
 ≤ 5 NSEC
- d) Fall time:
 ≤ 5 NSEC
- e) PRF: 60 Hz
- f) Jitter, stability:
 OK
- g) Prime power:
 120/280V
 50-60 Hz