


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-3*  
S.N.: *3233*  
Date: *SEPT 9 '86*



- a) Output signal amplitude:  
*0 TO +20 VOLTS*
- b) Pulse width:  
*5 TO 100 NSEC*
- c) Rise time:  
*≤ 200 PSEC*
- d) Fall time:  
*≤ 300 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 ± 1.0 V*  
*400 mA (MAX)*

*10 VOLTS/DIV*  
*2.0 NSEC/DIV*  
*500 kHz*