

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AV6-3-e-p-CA111

S.N.:

3279

Date:

Jul 29 86



a) Output signal amplitude:

0 TO +500 VOLTS

b) Pulse width:

≤ 5 NSEC (1/2 power)

c) Rise time:

NA

d) Fall time:

NR

e) PRF:

0 TO 20 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

ca 100 VOLTS/DIV

5 NSEC/DIV

$R_L = 50 \Omega$

10 KHz