


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MR-02-C-TR-P*

S.N.: *3287*

Date: *OCT 21 '66*



Ⓐ *1 nsec/div*
32 volts/div



Ⓑ *50 nsec/div*
10 volts/div
R_L = 50 Ω
50 MHz scope

- a) Output signal amplitude:
0 to +50 volts
- b) Pulse width:
20 to 200 nsec
- c) Rise time:
≤ 500 psec
- d) Fall time:
≤ 3 nsec
- e) PRF:
0 to 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Signature]