


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *QMM-2-P-PN-M-OT*  
S.N.: *3288*  
Date: *NOV 4 1966*



- a) Output signal amplitude:  
*0 to ±10 VOLTS*
- b) Pulse width:  
*5 to 100 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 to 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 ± 1.0 VOLTS*  
*300 mA (MAX)*

*10 VOLTS/DIV*  
*2 nSEC/DIV*


*h) OFFSET:*  
*0 to ±5 VOLTS*  
*to 50 Ω*

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PMMP-2-P-PN-M-OT-C*

S.N.: *3288 (MOD)*

Date: *FEB 20 87*



*40 dB ATTEN*  
*10 VOLTS/DIV*  
*1.0 MHz*

- a) Output signal amplitude:  
*0 TO ± 10 VOLTS*
- b) Pulse width:  
*5 TO 100 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz.*