


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-107C-AN-PN-N3*


S.N.: *3302*

Date: *Aug 29 '86*



Ⓐ *P_{out}*
R_L = 3.3 Ω
20 VOLTS/DIV
200 NSER/DIV
50 V/μA SLOPE

- a) Output signal amplitude:
0 TO ± 15 AmPS
- b) Pulse width: *50 VOLTS MAX*
50 NSER TO 1.0 usec
- c) Rise time:
≤ 10 NSER
- d) Fall time:
≤ 10 NSER
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:



Ⓑ *N_{out}*
(VS A)

- g) Prime power:
OK
120/240 V
50-60 Hz

[Signature]