


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-3-C-P-6A*

S.N.: *3346*

Date: *NOV 14 '86*



- a) Output signal amplitude:
0 TO +500 VOLTS
- b) Pulse width:
≤ 5 NSEC (100% POWER)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz

320 VOLTS/DIV
5.0 NSEC/DIV
20 KHz