


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P*  
S.N.: *3379*  
Date: *SEPT 2 1966*



- a) Output signal amplitude:  
*+25 TO +100 AMPS*
- b) Pulse width:  
*25 ± 5 NSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHZ.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*260 AMPS DIV.*  
*10 NSEC DIV*  
*500 KHZ*

*+15 ± 10 VOLTS*  
*200 mA (MAX)*  
*[Signature]*