


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-3-C-PN-UT1*

S.N.: *3389*

Date: *DEC 10 86*



70 db ATTEN
∴ ≈ 320 VOLTS/DIV
20 KHz

- a) Output signal amplitude:
0 TO ± 500 VOLTS
- b) Pulse width:
≤ 5 NS (HALF POWER)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz