


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-V-MV3-C-M-N*

S.N.: *3491*

Date: *MAR 10 87*



Ⓐ *60 dB ATTEN*
100 V SWT 10W
5 KHz



Ⓑ *AS Ⓐ BUT. 1NS/10W*
(RISE TIME)

- a) Output signal amplitude:
0 TO -100 V SWT3
- b) Pulse width:
2 TO 100 NS OR
- c) Rise time:
≤ 200 PS OR
- d) Fall time:
≤ 2 NS OR
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz